

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1264

Name: RITTER 1  
Sterilizer ID: V2178198  
10/14/2021 08:31 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 111.4 F 0.0  
2:00 114.4 F 0.0  
4:00 141.1 F 0.0  
6:00 173.6 F 2.1  
8:00 197.6 F 6.0  
10:00 211.6 F 10.8  
12:00 219.8 F 4.8  
14:00 233.3 F 10.1  
16:00 234.3 F 8.3  
18:00 244.5 F 13.0  
20:00 242.6 F 11.5  
22:00 255.6 F 18.2  
24:00 267.0 F 25.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.2  
1:00 272.6 F 28.9  
2:00 271.6 F 28.3  
3:00 272.7 F 28.9  
4:00 271.6 F 28.1

Min 271.5 F 28.1  
Max 273.0 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.8 F 0.0  
5:00 218.0 F 0.0  
10:00 196.4 F 0.0  
15:00 198.7 F 0.0  
20:00 197.1 F 0.0  
Dry Cycle Aborted

FILLING: 0:59  
HEATING: 25:27  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 22:43  
TOTAL CYCLE: 00:56:14

10/14/2021 09:27 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02