

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 001901  
11:30 26/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:44  
PRESSURIZING 1:53  
STERILIZING 2:47  
136.0 C 221kPa 2:47  
136.2 C 225kPa 3:17  
136.4 C 226kPa 3:47  
136.4 C 227kPa 4:17  
136.4 C 226kPa 4:47  
136.5 C 227kPa 5:17  
136.4 C 226kPa 5:47  
136.3 C 225kPa 6:17  
136.3 C 225kPa 6:17  
VENTING 6:17  
AIR DRYING 7:11  
CYCLE COMPLETE 28:27  
-----

Digital Signature #  
00A7ABDDAF4B1B03  
-----