
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000437
10:08 28/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:28
PRESSURIZING 2:35
STERILIZING 3:41
136.3 C 223kPa 3:41
136.1 C 230kPa 4:11
136.4 C 233kPa 4:41
136.1 C 230kPa 5:11
136.0 C 230kPa 5:41
136.3 C 233kPa 6:11
136.0 C 231kPa 6:41
136.4 C 233kPa 7:11
136.4 C 233kPa 7:11
VENTING 7:11
AIR DRYING 8:03
CYCLE COMPLETE 8:58

Digital Signature #
435C3997AFFE4BA4
