

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001556  
11:09 26/04/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:07  
PRESSURIZING 3:09  
STERILIZING 4:15  
136.4 C 225kPa 4:15  
136.4 C 225kPa 4:45  
136.4 C 224kPa 5:15  
136.4 C 224kPa 5:45  
136.2 C 221kPa 6:15  
136.4 C 224kPa 6:45  
136.4 C 223kPa 7:15  
136.4 C 224kPa 7:45  
136.4 C 224kPa 7:45  
VENTING 7:45  
AIR DRYING 8:35  
CYCLE COMPLETE 8:43

-----  
Digital Signature #  
B94B8F5AA912494E  
-----