
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001395
6:23 09/08/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:19
PRESSURIZING 4:58
STERILIZING 6:20
136.4 C 220kPa 6:20
136.3 C 221kPa 6:50
136.4 C 221kPa 7:20
136.4 C 216kPa 7:50
136.4 C 223kPa 8:20
136.1 C 220kPa 8:50
136.4 C 223kPa 9:20
136.4 C 222kPa 9:50
VENTING 9:50
AIR DRYING 10:52
CYCLE COMPLETE 11:14

Digital Signature #
133B586D9803A0CD
