
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 003991
13:36 11/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:20
PRESSURIZING 2:28
STERILIZING 3:32
136.4 C 220kPa 3:32
136.2 C 223kPa 4:02
136.4 C 222kPa 4:32
136.4 C 225kPa 5:02
136.4 C 224kPa 5:32
136.3 C 223kPa 6:02
136.4 C 225kPa 6:32
136.2 C 222kPa 7:02
136.2 C 222kPa 7:02
VENTING 7:02
AIR DRYING 9:01
CYCLE COMPLETE 9:25

Digital Signature #
625F04FB63A6F1C5
