
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000510
10:50 21/05/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:08
PRESSURIZING 3:13
STERILIZING 4:27
136.4 C 220kPa 4:27
136.4 C 232kPa 4:57
136.2 C 230kPa 5:27
136.4 C 231kPa 5:57
136.2 C 231kPa 6:27
136.2 C 232kPa 6:57
136.3 C 231kPa 7:27
136.4 C 233kPa 7:57
136.4 C 233kPa 7:57
VENTING 7:57
AIR DRYING 8:51
CYCLE COMPLETE 23:51

Digital Signature #
39E8B95204F9E34E
