

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4805

Name: RITTER 1  
Sterilizer ID: V2178198  
08/12/2024 02:41 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 104.2 F 0.0  
2:00 109.6 F 0.3  
4:00 137.9 F 1.0  
6:00 168.7 F 1.4  
8:00 187.3 F 4.3  
10:00 203.8 F 7.3  
12:00 214.4 F 10.8  
14:00 224.0 F 7.6  
16:00 228.1 F 8.6  
18:00 238.4 F 12.9  
20:00 236.9 F 9.8  
22:00 238.3 F 9.5  
24:00 246.7 F 13.6  
26:00 257.7 F 19.5  
28:00 267.5 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.4  
1:00 271.8 F 28.8  
2:00 272.6 F 29.2  
3:00 272.5 F 29.0  
4:00 272.2 F 28.7

Min 271.2 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.7 F 0.0  
5:00 199.4 F 0.0  
10:00 209.4 F 0.0  
15:00 199.9 F 0.0  
20:00 213.9 F 0.0  
25:00 212.3 F 0.0  
30:00 213.6 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 29:37  
STERILIZING: 4:00  
VENTING: 3:52  
DRYING: 30:00  
TOTAL CYCLE: 01:08:19  
CYCLE COMPLETE

08/12/2024 03:49 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_