
Midmark M11 - v1.0.0
Total Cycles: 4518

Name: RITTER 1
Sterilizer ID: V2178198
05/02/2024 01:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.1 F 0.0
2:00 140.0 F 2.0
4:00 153.0 F 5.3
6:00 184.9 F 9.3
8:00 209.5 F 10.2
10:00 218.5 F 7.7
12:00 228.5 F 12.7
14:00 232.9 F 9.0
16:00 239.1 F 10.2
18:00 244.8 F 13.0
20:00 257.1 F 19.4
22:00 268.1 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.5
1:00 272.3 F 29.3
2:00 271.8 F 29.3
3:00 271.7 F 29.0
4:00 272.1 F 29.0

Min 270.9 F 28.4
Max 273.1 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.7 F 0.0
5:00 206.8 F 0.0
10:00 223.5 F 0.0
15:00 234.4 F 0.0
20:00 225.9 F 0.0
25:00 222.8 F 0.0
30:00 224.7 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 23:21
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:02:03
CYCLE COMPLETE

05/02/2024 02:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____