
Midmark M11 - v1.0.0
Total Cycles: 3134

Name: RITTER 1
Sterilizer ID: V2178198
04/20/2023 08:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.5 F 0.0
2:00 177.2 F 0.0
4:00 186.1 F 0.0
6:00 209.0 F 3.4
8:00 224.7 F 8.3
10:00 239.0 F 13.1
12:00 235.8 F 9.8
14:00 238.8 F 8.8
16:00 245.0 F 12.1
18:00 246.2 F 13.0
20:00 258.9 F 19.9
22:00 270.1 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 27.9
1:00 272.6 F 28.5
2:00 271.5 F 27.9
3:00 272.7 F 28.7
4:00 271.8 F 28.0

Min 271.4 F 27.8
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 213.4 F 0.0
10:00 229.9 F 0.0
15:00 229.0 F 0.0
20:00 227.0 F 0.0
25:00 226.2 F 0.0
30:00 224.8 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 22:54
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:55
CYCLE COMPLETE

04/20/2023 09:33 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____