
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001623
11:25 11/26/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:18
PRESSURIZING 7:08
STERILIZING 8:49
136.0 C 220kPa 8:49
136.0 C 218kPa 9:19
136.4 C 223kPa 9:49
136.2 C 220kPa 10:19
135.9 C 219kPa 10:49
136.4 C 219kPa 11:19
136.4 C 222kPa 11:49
136.2 C 221kPa 12:19
VENTING 12:19
AIR DRYING 13:21
CYCLE COMPLETE 43:21

Digital Signature #
C623D6ACF8703680
