

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 007919  
15:11 02/10/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:21  
PRESSURIZING 2:25  
STERILIZING 3:28  
136.0 C 220kPa 3:28  
136.0 C 222kPa 3:59  
136.4 C 224kPa 4:29  
136.0 C 221kPa 4:59  
136.2 C 220kPa 5:29  
136.4 C 224kPa 5:59  
136.4 C 223kPa 6:29  
136.4 C 224kPa 6:59  
136.4 C 224kPa 6:59  
VENTING 6:59  
AIR DRYING 7:49  
CYCLE COMPLETE 37:02

-----  
Digital Signature #  
C2750F5C850F0B45  
-----