
Midmark M11 - v1.0.5
Total Cycles: 912

Name: RITTER2
Sterilizer ID: V2313228
03/28/2022 08:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 128.4 F 0.0
2:00 128.1 F 0.0
4:00 140.8 F 0.2
6:00 183.2 F 1.7
8:00 204.7 F 6.5
10:00 223.1 F 11.4
12:00 229.2 F 7.5
14:00 239.9 F 13.2
16:00 240.2 F 11.6
18:00 240.3 F 11.0
20:00 254.4 F 18.2
22:00 266.5 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.7 F 29.9
2:00 272.2 F 29.3
3:00 271.4 F 29.1
4:00 272.6 F 29.9

Min 271.4 F 28.9
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.7 F 0.0
5:00 208.3 F 0.0
10:00 210.9 F 0.0
15:00 212.0 F 0.0
20:00 209.9 F 0.0
25:00 211.1 F 0.0
30:00 212.6 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 23:36
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:01:38
CYCLE COMPLETE

03/28/2022 09:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____