

-----  
Midmark M11 - v1.0.0  
Total Cycles: 635

Name: MIDMARK  
Sterilizer ID: V2178198  
01/27/2021 11:22 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 146.7 F 0.0  
2:00 146.8 F 1.6  
4:00 162.1 F 5.1  
6:00 193.3 F 10.3  
8:00 219.6 F 4.9  
10:00 226.8 F 11.4  
12:00 233.1 F 8.0  
14:00 240.4 F 10.4  
16:00 245.8 F 13.0  
18:00 259.0 F 20.0  
20:00 270.3 F 27.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.1  
1:00 272.3 F 28.9  
2:00 272.5 F 28.8  
3:00 272.2 F 28.7  
4:00 272.3 F 28.7  
  
Min 270.7 F 28.0  
Max 273.0 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.3 F 0.0  
5:00 241.8 F 0.0  
10:00 226.5 F 0.0  
15:00 228.7 F 0.0  
20:00 220.7 F 0.0  
25:00 212.4 F 0.0  
30:00 207.5 F 0.0  
Drying Complete

FILLING: 1:10  
HEATING: 20:39  
STERILIZING: 4:00  
VENTING: 2:56  
DRYING: 30:00  
TOTAL CYCLE: 00:58:46  
CYCLE COMPLETE

01/27/2021 12:20 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_