

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2191

Name: RITTER2  
Sterilizer ID: V2313228  
03/29/2023 08:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.5 F 0.0  
2:00 174.2 F 3.0  
4:00 173.6 F 6.7  
6:00 196.2 F 11.9  
8:00 217.5 F 6.7  
10:00 234.0 F 12.2  
12:00 234.6 F 9.6  
14:00 238.3 F 9.0  
16:00 248.7 F 15.1  
18:00 261.3 F 22.3  
20:00 272.6 F 29.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.2  
1:00 272.4 F 29.7  
2:00 271.1 F 29.1  
3:00 272.5 F 29.9  
4:00 271.5 F 29.3  
  
Min 271.1 F 29.1  
Max 272.6 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.8 F 0.0  
5:00 212.7 F 0.0  
10:00 223.0 F 0.0  
15:00 230.1 F 0.0  
20:00 236.2 F 0.0  
25:00 216.2 F 0.0  
30:00 218.2 F 0.0  
Drying Complete

FILLING: 1:12  
HEATING: 20:27  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:58:52  
CYCLE COMPLETE

03/29/2023 09:53 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_