
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000648
7:11 13/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:20
PRESSURIZING 2:27
STERILIZING 3:31
136.2 C 228kPa 3:31
136.4 C 230kPa 4:01
136.1 C 230kPa 4:31
136.4 C 232kPa 5:01
136.4 C 232kPa 5:31
136.2 C 230kPa 6:01
136.4 C 232kPa 6:31
135.8 C 228kPa 7:01
135.8 C 228kPa 7:01
VENTING 7:01
AIR DRYING 7:51
CYCLE COMPLETE 7:54

Digital Signature #
5653BD5975F66728
