
Midmark M11 - v1.0.0
Total Cycles: 1742

Name: RITTER 1
Sterilizer ID: V2178198
03/24/2022 12:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.9 F 0.0
2:00 177.9 F 3.1
4:00 194.0 F 7.4
6:00 229.2 F 13.2
8:00 231.3 F 8.2
10:00 242.9 F 13.5
12:00 240.8 F 11.1
14:00 237.9 F 9.4
16:00 252.3 F 16.4
18:00 264.3 F 23.4
20:00 271.7 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.2 F 28.6
2:00 272.5 F 28.9
3:00 272.7 F 29.2
4:00 271.8 F 28.9

Min 271.6 F 28.5
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.3 F 0.0
5:00 229.3 F 0.0
10:00 225.5 F 0.0
15:00 223.8 F 0.0
20:00 188.3 F 0.0
25:00 181.6 F 0.0
30:00 179.1 F 0.0
Drying Complete

FILLING: 1:47
HEATING: 20:30
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:27
CYCLE COMPLETE

03/24/2022 01:55 PM
ACCEPTED: X
REJECTED:
Operator: U02