
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001609
9:48 11/24/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:26
PRESSURIZING 6:08
STERILIZING 7:43
136.0 C 220kPa 7:43
136.4 C 222kPa 8:13
136.3 C 221kPa 8:43
136.4 C 222kPa 9:13
135.9 C 219kPa 9:43
136.3 C 221kPa 10:14
136.4 C 221kPa 10:44
136.0 C 218kPa 11:13
VENTING 11:13
AIR DRYING 12:18
CYCLE COMPLETE 28:19

Digital Signature #
DEBC9007FA01DDA9
