
Midmark M11 - v1.0.5
Total Cycles: 2890

Name: RITTER2
Sterilizer ID: V2313228
09/15/2023 11:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 139.5 F 0.0
2:00 139.1 F 0.0
4:00 156.6 F 1.1
6:00 180.2 F 4.7
8:00 197.6 F 8.8
10:00 215.9 F 13.3
12:00 224.7 F 8.3
14:00 237.4 F 13.2
16:00 238.9 F 10.5
18:00 236.1 F 8.5
20:00 249.3 F 15.2
22:00 260.5 F 21.6
24:00 271.0 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 273.2 F 29.7
2:00 271.6 F 29.5
3:00 272.6 F 29.6
4:00 271.4 F 29.3

Min 271.3 F 29.1
Max 273.3 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.3 F 0.0
5:00 229.8 F 0.0
10:00 184.4 F 0.0
15:00 179.6 F 0.0
20:00 174.2 F 0.0
25:00 172.2 F 0.0
30:00 171.1 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 25:28
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:03:56
CYCLE COMPLETE

09/15/2023 12:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____