
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000481
15:04 17/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:53
PRESSURIZING 2:56
STERILIZING 3:51
136.2 C 221kPa 3:51
136.2 C 223kPa 4:21
136.4 C 223kPa 4:51
136.2 C 224kPa 5:21
136.4 C 222kPa 5:51
135.9 C 219kPa 6:21
136.3 C 223kPa 6:51
136.4 C 223kPa 7:21
136.4 C 223kPa 7:21
VENTING 7:21
AIR DRYING 8:13
CYCLE COMPLETE 8:23

Digital Signature #
57F691EA81FB1164
