
Midmark M11 - v1.0.5
Total Cycles: 2665

Name: RITTER2
Sterilizer ID: V2313228
07/19/2023 08:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.8 F 0.0
2:00 72.1 F 0.0
4:00 78.7 F 0.0
6:00 103.8 F 0.0
8:00 148.2 F 0.2
10:00 166.9 F 2.0
12:00 184.4 F 5.1
14:00 212.8 F 8.3
16:00 219.5 F 12.0
18:00 222.2 F 5.9
20:00 233.4 F 10.1
22:00 235.7 F 9.6
24:00 241.3 F 11.9
26:00 239.2 F 10.3
28:00 251.4 F 16.3
30:00 261.8 F 22.3
32:00 271.5 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.5 F 29.4
2:00 271.6 F 28.6
3:00 272.2 F 29.6
4:00 272.1 F 29.1

Min 271.2 F 28.6
Max 272.5 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.9 F 0.0
5:00 231.8 F 0.0
10:00 221.7 F 0.0
15:00 216.5 F 0.0
20:00 212.9 F 0.0
25:00 216.5 F 0.0
30:00 216.4 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 32:41
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:10:47
CYCLE COMPLETE

07/19/2023 09:14 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____