
Midmark M11 - v1.0.0
Total Cycles: 3791

Name: RITTER 1
Sterilizer ID: V2178198
09/18/2023 03:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.1 F 0.0
2:00 176.1 F 0.8
4:00 205.0 F 3.9
6:00 211.7 F 7.3
8:00 222.9 F 11.2
10:00 226.7 F 5.8
12:00 238.4 F 11.0
14:00 234.6 F 8.3
16:00 243.3 F 11.5
18:00 246.5 F 13.3
20:00 257.7 F 19.3
22:00 268.0 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.6 F 28.7
2:00 271.3 F 28.0
3:00 272.4 F 28.9
4:00 271.8 F 28.3

Min 271.2 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.3 F 0.0
5:00 230.6 F 0.0
10:00 208.0 F 0.0
15:00 196.2 F 0.0
20:00 175.4 F 0.0
25:00 171.9 F 0.0
30:00 171.1 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 23:21
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:02:27
CYCLE COMPLETE

09/18/2023 04:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____