
Midmark M11 - v1.0.0
Total Cycles: 1826

Name: RITTER 1
Sterilizer ID: V2178198
04/14/2022 12:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.0 F 0.0
2:00 179.9 F 3.4
4:00 197.1 F 7.5
6:00 222.0 F 12.9
8:00 226.2 F 7.2
10:00 237.3 F 12.3
12:00 236.2 F 9.3
14:00 237.6 F 8.9
16:00 248.1 F 14.4
18:00 260.8 F 21.4
20:00 272.2 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.4 F 29.0
2:00 271.2 F 28.5
3:00 272.4 F 29.0
4:00 271.2 F 28.3

Min 271.1 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.2 F 0.0
5:00 229.6 F 0.0
10:00 227.3 F 0.0
15:00 231.7 F 0.0
20:00 230.7 F 0.0
25:00 230.7 F 0.0
30:00 226.0 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 20:29
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:34
CYCLE COMPLETE

04/14/2022 01:42 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____