
Midmark M11 - v1.0.5
Total Cycles: 1006

Name: RITTER2
Sterilizer ID: V2313228
04/19/2022 01:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.7 F 0.0
2:00 182.6 F 0.8
4:00 195.5 F 4.4
6:00 211.6 F 9.6
8:00 225.5 F 7.9
10:00 235.5 F 11.8
12:00 237.1 F 10.3
14:00 237.0 F 9.5
16:00 251.5 F 16.8
18:00 263.8 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.6 F 29.8
2:00 271.7 F 29.5
3:00 272.5 F 29.7
4:00 271.5 F 29.2

Min 271.5 F 29.1
Max 272.6 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.4 F 0.0
5:00 203.8 F 0.0
10:00 219.2 F 0.0
15:00 220.7 F 0.0
20:00 221.0 F 0.0
25:00 221.2 F 0.0
30:00 220.9 F 0.0
Drying Complete

FILLING: 2:00
HEATING: 19:58
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:14
CYCLE COMPLETE

04/19/2022 02:16 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____