
Midmark M11 - v1.0.5
Total Cycles: 2361

Name: RITTER2
Sterilizer ID: V2313228
05/12/2023 12:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 133.6 F 0.0
2:00 135.0 F 1.2
4:00 155.6 F 4.2
6:00 180.5 F 8.3
8:00 199.8 F 13.0
10:00 217.7 F 6.6
12:00 230.5 F 11.4
14:00 231.2 F 8.2
16:00 242.5 F 12.1
18:00 243.5 F 12.6
20:00 255.8 F 19.0
22:00 266.8 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.3 F 29.9
2:00 272.1 F 29.3
3:00 271.2 F 28.9
4:00 272.0 F 29.8

Min 271.1 F 28.8
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.4 F 0.0
5:00 212.7 F 0.0
10:00 217.4 F 0.0
15:00 215.3 F 0.0
20:00 216.4 F 0.0
25:00 218.0 F 0.0
30:00 216.0 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 23:36
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:01:45
CYCLE COMPLETE

05/12/2023 01:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____