

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2203

Name: RITTER2  
Sterilizer ID: V2313228  
03/31/2023 08:52 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 169.8 F 0.0  
2:00 166.5 F 2.2  
4:00 173.0 F 6.4  
6:00 198.0 F 12.1  
8:00 222.0 F 7.1  
10:00 237.7 F 11.6  
12:00 242.5 F 12.4  
14:00 242.7 F 11.8  
16:00 257.2 F 19.4  
18:00 269.5 F 27.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.4 F 28.9  
2:00 272.7 F 29.5  
3:00 271.8 F 28.5  
4:00 272.9 F 29.4  
  
Min 271.5 F 28.4  
Max 273.0 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 224.4 F 0.0  
Dry Cycle Aborted

FILLING: 1:28  
HEATING: 18:54  
STERILIZING: 4:00  
VENTING: 2:48  
DRYING: 2:21  
TOTAL CYCLE: 00:29:31

03/31/2023 09:21 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_