
Midmark M11 - v1.0.5
Total Cycles: 3738

Name: RITTER2
Sterilizer ID: V2313228
06/26/2024 03:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 114.7 F 0.0
2:00 117.6 F 1.2
4:00 135.6 F 3.8
6:00 163.7 F 7.1
8:00 187.0 F 10.8
10:00 204.0 F 4.8
12:00 216.4 F 6.0
14:00 227.9 F 10.0
16:00 232.3 F 9.5
18:00 239.9 F 11.7
20:00 238.7 F 10.3
22:00 251.0 F 16.3
24:00 261.8 F 22.5
26:00 271.9 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.5 F 29.6
2:00 271.4 F 28.9
3:00 272.4 F 29.7
4:00 271.8 F 29.1

Min 271.3 F 28.9
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.2 F 0.0
5:00 214.6 F 0.0
10:00 181.9 F 0.0
15:00 200.2 F 0.0
20:00 198.4 F 0.0
25:00 199.3 F 0.0
30:00 190.7 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 26:33
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:05:26
CYCLE COMPLETE

06/26/2024 04:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____