
Midmark M11 - v1.0.0
Total Cycles: 2278

Name: RITTER 1
Sterilizer ID: V2178198
08/31/2022 06:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.8 F 0.0
2:00 199.6 F 1.3
4:00 215.3 F 6.5
6:00 233.9 F 11.9
8:00 231.7 F 8.3
10:00 241.8 F 11.3
12:00 244.2 F 12.7
14:00 243.3 F 11.9
16:00 256.7 F 18.9
18:00 268.4 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.6 F 28.9
2:00 271.7 F 28.5
3:00 272.6 F 28.9
4:00 271.6 F 28.3

Min 271.5 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.2 F 0.0
5:00 238.1 F 0.0
10:00 229.5 F 0.0
15:00 230.2 F 0.0
20:00 227.6 F 0.0
25:00 229.4 F 0.0
30:00 230.6 F 0.0

Drying Complete

FILLING: 1:57
HEATING: 19:12
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:23
CYCLE COMPLETE

08/31/2022 07:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____