
Midmark M11 - v1.0.0
Total Cycles: 4931

Name: RITTER 1
Sterilizer ID: V2178198
09/23/2024 01:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 88.3 F 0.0
2:00 89.5 F 0.0
4:00 115.4 F 0.0
6:00 145.8 F 0.1
8:00 172.0 F 0.2
10:00 194.5 F 0.5
12:00 208.9 F 3.2
14:00 218.5 F 6.2
16:00 230.6 F 9.6
18:00 238.5 F 12.0
20:00 239.1 F 10.6
22:00 235.8 F 8.4
24:00 242.9 F 11.3
26:00 247.0 F 13.5
28:00 258.3 F 19.5
30:00 268.4 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 271.9 F 28.4
2:00 272.4 F 28.8
3:00 272.7 F 28.8
4:00 272.6 F 28.7

Min 271.5 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.6 F 0.0
5:00 191.3 F 0.0
10:00 190.2 F 0.0
15:00 187.8 F 0.0
20:00 194.6 F 0.0
25:00 210.5 F 0.0
30:00 213.0 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 31:30
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:10:04
CYCLE COMPLETE

09/23/2024 02:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____