

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002535  
8:55 24/02/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CYCLE ABORTED 0:56  
STOP BUTTON PRESSED  
NOT STERILE  
END TIME 0:56  
-----

Digital Signature #  
DE93B3B1AA5B1D8F  
-----