
Midmark M11 - v1.0.5
Total Cycles: 1271

Name: RITTER2
Sterilizer ID: V2313228
07/08/2022 06:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.7 F 0.0
2:00 74.0 F 0.0
4:00 88.5 F 0.0
6:00 120.5 F 0.0
8:00 155.9 F 0.2
10:00 182.7 F 0.7
12:00 195.3 F 3.9
14:00 212.4 F 7.1
16:00 227.9 F 10.6
18:00 230.8 F 8.4
20:00 236.1 F 10.0
22:00 238.1 F 9.6
24:00 244.0 F 12.9
26:00 242.5 F 11.7
28:00 254.5 F 17.9
30:00 265.0 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.8
1:00 272.2 F 29.4
2:00 272.1 F 28.8
3:00 272.1 F 29.4
4:00 272.5 F 29.1

Min 271.2 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.1 F 0.0
5:00 236.0 F 0.0
10:00 224.6 F 0.0
15:00 229.7 F 0.0
20:00 214.5 F 0.0
25:00 233.8 F 0.0
30:00 168.6 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 31:49
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:09:59
CYCLE COMPLETE

07/08/2022 08:04 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____