
Midmark M11 - v1.0.0
Total Cycles: 3249

Name: RITTER 1
Sterilizer ID: V2178198
05/17/2023 10:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.8 F 0.0
2:00 184.9 F 1.6
4:00 198.0 F 6.4
6:00 225.8 F 11.8
8:00 228.5 F 7.4
10:00 240.9 F 13.0
12:00 239.7 F 10.8
14:00 237.9 F 9.3
16:00 252.3 F 16.3
18:00 264.2 F 23.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.6 F 28.9
2:00 271.8 F 28.2
3:00 272.0 F 28.7
4:00 272.5 F 28.8

Min 271.3 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.7 F 0.0
5:00 232.0 F 0.0
10:00 226.5 F 0.0
15:00 226.8 F 0.0
20:00 224.5 F 0.0
25:00 223.0 F 0.0
30:00 221.9 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 19:59
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:54
CYCLE COMPLETE

05/17/2023 11:00 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____