
Midmark M11 - v1.0.0
Total Cycles: 5446

Name: RITTER 1
Sterilizer ID: V2178198
03/26/2025 08:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.5 F 0.0
2:00 74.0 F 0.4
4:00 92.5 F 1.9
6:00 127.0 F 4.4
8:00 158.7 F 7.1
10:00 176.0 F 10.6
12:00 186.0 F 5.0
14:00 205.6 F 4.9
16:00 217.1 F 8.3
18:00 228.7 F 12.2
20:00 228.7 F 7.1
22:00 240.7 F 12.1
24:00 237.8 F 9.6
26:00 250.8 F 15.8
28:00 261.7 F 22.0
30:00 272.0 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.8 F 29.1
2:00 271.7 F 28.5
3:00 272.9 F 29.1
4:00 271.8 F 28.3

Min 271.6 F 28.3
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.9 F 0.0
5:00 184.7 F 0.0
10:00 192.4 F 0.0
15:00 198.4 F 0.0
20:00 203.8 F 0.0
25:00 205.8 F 0.0
30:00 210.2 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 30:33
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:09:00
CYCLE COMPLETE

03/26/2025 09:42 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____