
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000856
11:02 10/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:31
PRESSURIZING 3:35
STERILIZING 4:51
136.0 C 228kPa 4:51
136.4 C 232kPa 5:21
136.4 C 230kPa 5:51
136.0 C 229kPa 6:21
136.4 C 232kPa 6:51
136.3 C 231kPa 7:21
136.4 C 231kPa 7:51
136.2 C 230kPa 8:21
136.2 C 230kPa 8:21
VENTING 8:21
AIR DRYING 9:14
CYCLE COMPLETE 24:14

Digital Signature #
81A4629FD5CB87E6
