
Midmark M11 - v1.0.0
Total Cycles: 5100

Name: RITTER 1
Sterilizer ID: V2178198
11/15/2024 04:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.9 F 0.0
2:00 192.4 F 0.5
4:00 205.2 F 3.0
6:00 221.2 F 7.7
8:00 232.7 F 12.6
10:00 233.4 F 8.9
12:00 240.1 F 10.3
14:00 244.8 F 12.9
16:00 243.5 F 11.9
18:00 256.4 F 18.6
20:00 267.7 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.2
1:00 272.5 F 29.0
2:00 271.5 F 28.1
3:00 272.5 F 29.0
4:00 271.6 F 28.2

Min 271.3 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.4 F 0.0
5:00 214.9 F 0.0
10:00 202.9 F 0.0
15:00 221.2 F 0.0
20:00 232.3 F 0.0
25:00 230.1 F 0.0
30:00 229.2 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 21:13
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:00:15
CYCLE COMPLETE

11/15/2024 05:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____