

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3432

Name: RITTER2  
Sterilizer ID: V2313228  
03/04/2024 02:50 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 94.3 F 0.0  
2:00 110.8 F 0.0  
4:00 118.5 F 0.9  
6:00 143.4 F 3.6  
8:00 173.3 F 7.4  
10:00 194.0 F 11.5  
12:00 210.5 F 4.2  
14:00 218.2 F 8.5  
16:00 231.8 F 12.9  
18:00 234.8 F 9.6  
20:00 238.4 F 9.6  
22:00 246.4 F 13.9  
24:00 258.3 F 20.4  
26:00 269.1 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.0  
1:00 272.5 F 29.6  
2:00 271.4 F 28.9  
3:00 272.5 F 29.7  
4:00 271.8 F 29.0

Min 271.3 F 28.8  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.0 F 0.0  
5:00 221.6 F 0.0  
10:00 217.1 F 0.0  
15:00 208.6 F 0.0  
20:00 208.7 F 0.0  
25:00 213.7 F 0.0  
30:00 211.8 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 27:06  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 01:05:22  
CYCLE COMPLETE

03/04/2024 03:55 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_