
Midmark M11 - v1.0.5
Total Cycles: 1024

Name: RITTER2
Sterilizer ID: V2313228
04/25/2022 08:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 137.8 F 0.0
2:00 136.8 F 1.4
4:00 142.9 F 4.0
6:00 165.0 F 8.0
8:00 193.5 F 12.8
10:00 214.4 F 6.5
12:00 225.4 F 11.6
14:00 230.7 F 8.3
16:00 241.8 F 12.0
18:00 243.4 F 12.6
20:00 255.1 F 18.9
22:00 266.4 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.3
1:00 272.4 F 29.9
2:00 271.4 F 29.1
3:00 272.1 F 30.1
4:00 272.2 F 29.6

Min 271.0 F 29.1
Max 272.6 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.8 F 0.0
5:00 204.1 F 0.0
10:00 212.6 F 0.0
15:00 220.1 F 0.0
20:00 222.8 F 0.0
25:00 229.2 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 23:38
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:01:46
CYCLE COMPLETE

04/25/2022 09:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____