

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3426

Name: RITTER2  
Sterilizer ID: V2313228  
03/01/2024 11:19 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 157.2 F 0.0  
2:00 153.5 F 1.1  
4:00 160.5 F 3.8  
6:00 184.3 F 7.8  
8:00 202.8 F 12.7  
10:00 224.0 F 6.8  
12:00 233.4 F 11.6  
14:00 232.5 F 8.4  
16:00 241.8 F 11.5  
18:00 244.8 F 12.7  
20:00 256.8 F 19.1  
22:00 267.3 F 25.8  
24:00 271.6 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.1  
1:00 272.7 F 29.4  
2:00 271.9 F 29.5  
3:00 272.4 F 29.3  
4:00 271.8 F 29.5

Min 271.4 F 28.9  
Max 273.2 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 171.6 F 0.0  
5:00 227.1 F 0.0  
10:00 222.2 F 0.0  
15:00 194.4 F 0.0  
20:00 178.6 F 0.0  
25:00 175.1 F 0.0  
30:00 173.3 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 24:05  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:02:26  
CYCLE COMPLETE

03/01/2024 12:21 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_