
Midmark M11 - v1.0.0
Total Cycles: 5168

Name: RITTER 1
Sterilizer ID: V2178198
12/06/2024 11:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.2 F 0.0
2:00 150.2 F 0.1
4:00 157.8 F 0.1
6:00 190.0 F 0.5
8:00 211.3 F 0.9
10:00 220.0 F 3.6
12:00 230.8 F 7.2
14:00 239.1 F 11.2
16:00 233.3 F 7.6
18:00 246.4 F 13.4
20:00 243.8 F 11.9
22:00 241.3 F 10.7
24:00 253.8 F 16.9
26:00 264.1 F 23.0
28:00 272.8 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 271.9 F 28.3
2:00 272.4 F 28.7
3:00 272.8 F 28.8
4:00 272.6 F 28.6

Min 271.6 F 27.9
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.5 F 0.0
5:00 215.5 F 0.0
10:00 239.5 F 0.0
15:00 233.6 F 0.0
20:00 235.9 F 0.0
25:00 225.9 F 0.0
30:00 223.0 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 28:26
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:06:57
CYCLE COMPLETE

12/06/2024 12:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____