

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1923

Name: RITTER2  
Sterilizer ID: V2313228  
01/11/2023 04:17 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 174.1 F 0.0  
2:00 175.0 F 3.6  
4:00 190.4 F 8.3  
6:00 210.6 F 10.2  
8:00 222.8 F 9.2  
10:00 232.8 F 9.2  
12:00 242.6 F 13.1  
14:00 242.9 F 12.2  
16:00 255.9 F 19.1  
18:00 267.3 F 26.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.1  
1:00 272.4 F 29.8  
2:00 271.8 F 29.1  
3:00 271.8 F 29.7  
4:00 272.4 F 29.7  
  
Min 271.4 F 28.9  
Max 272.6 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.0 F 0.0  
5:00 235.5 F 0.0  
10:00 216.6 F 0.0  
15:00 211.0 F 0.0  
20:00 219.1 F 0.0  
25:00 220.4 F 0.0  
30:00 223.6 F 0.0  
Drying Complete

FILLING: 1:44  
HEATING: 19:24  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:58:18  
CYCLE COMPLETE

01/11/2023 05:15 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_