
Midmark M11 - v1.0.5
Total Cycles: 4340

Name: RITTER2
Sterilizer ID: V2313228
02/07/2025 09:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.5 F 0.0
2:00 78.3 F 0.0
4:00 80.6 F 0.3
6:00 107.9 F 2.6
8:00 146.2 F 5.5
10:00 167.3 F 8.7
12:00 187.0 F 12.5
14:00 209.0 F 3.6
16:00 221.3 F 7.0
18:00 231.5 F 10.7
20:00 230.1 F 7.8
22:00 238.9 F 11.5
24:00 236.4 F 9.2
26:00 248.8 F 15.3
28:00 259.7 F 21.4
30:00 270.0 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.1
1:00 272.9 F 29.7
2:00 271.8 F 29.2
3:00 272.8 F 29.7
4:00 271.8 F 29.1

Min 271.8 F 29.0
Max 273.1 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.3 F 0.0
5:00 178.8 F 0.0
10:00 183.4 F 0.0
15:00 187.5 F 0.0
20:00 191.8 F 0.0
25:00 199.1 F 0.0
30:00 201.1 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:01
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:09:19
CYCLE COMPLETE

02/07/2025 11:04 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____