

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3017

Name: RITTER2  
Sterilizer ID: V2313228  
10/17/2023 09:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 138.6 F 0.0  
2:00 142.6 F 1.7  
4:00 158.7 F 4.7  
6:00 182.9 F 9.1  
8:00 205.8 F 7.6  
10:00 218.5 F 8.2  
12:00 233.3 F 13.6  
14:00 237.5 F 10.7  
16:00 235.9 F 9.0  
18:00 250.6 F 15.8  
20:00 262.2 F 22.7  
22:00 272.9 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.3  
1:00 272.5 F 29.7  
2:00 272.3 F 30.0  
3:00 272.1 F 29.4  
4:00 272.5 F 30.1

Min 271.5 F 29.2  
Max 272.7 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.1 F 0.0  
5:00 225.3 F 0.0  
10:00 194.6 F 0.0  
15:00 191.4 F 0.0  
20:00 189.2 F 0.0  
25:00 187.1 F 0.0  
30:00 187.0 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 23:06  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:01:15  
CYCLE COMPLETE

10/17/2023 10:56 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_