
Midmark M11 - v1.0.5
Total Cycles: 1044

Name: RITTER2
Sterilizer ID: V2313228
04/28/2022 07:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.8 F 0.0
2:00 74.2 F 0.0
4:00 82.8 F 0.0
6:00 109.9 F 0.1
8:00 161.1 F 0.3
10:00 182.3 F 3.5
12:00 216.7 F 7.4
14:00 215.2 F 11.3
16:00 219.7 F 5.1
18:00 233.9 F 9.8
20:00 235.5 F 9.8
22:00 242.5 F 12.1
24:00 240.7 F 10.8
26:00 253.9 F 17.5
28:00 265.3 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.8 F 29.5
2:00 271.8 F 28.5
3:00 272.4 F 29.5
4:00 272.6 F 29.0

Min 271.4 F 28.5
Max 273.0 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.2 F 0.0
5:00 206.9 F 0.0
10:00 228.6 F 0.0
15:00 234.1 F 0.0
20:00 226.1 F 0.0
25:00 230.1 F 0.0
30:00 229.1 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 29:50
STERILIZING: 4:00
VENTING: 2:55
DRYING: 30:00
TOTAL CYCLE: 01:07:33
CYCLE COMPLETE

04/28/2022 08:23 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____