
Midmark M11 - v1.0.5
Total Cycles: 3287

Name: RITTER2
Sterilizer ID: V2313228
01/11/2024 09:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.6 F 0.0
2:00 80.3 F 0.0
4:00 133.1 F 0.0
6:00 106.2 F 1.1
8:00 140.6 F 4.1
10:00 172.0 F 7.4
12:00 189.2 F 11.8
14:00 217.1 F 5.2
16:00 227.0 F 9.4
18:00 236.5 F 11.0
20:00 237.1 F 10.1
22:00 234.7 F 8.5
24:00 246.4 F 13.9
26:00 258.3 F 20.4
28:00 269.0 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.2 F 29.1
2:00 272.4 F 29.6
3:00 271.5 F 29.4
4:00 272.0 F 29.1

Min 271.1 F 28.9
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.0 F 0.0
5:00 232.6 F 0.0
10:00 191.3 F 0.0
15:00 182.1 F 0.0
20:00 177.8 F 0.0
25:00 176.6 F 0.0
30:00 176.9 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 29:04
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 01:06:56
CYCLE COMPLETE

01/11/2024 10:15 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____