

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2978

Name: RITTER2  
Sterilizer ID: V2313228  
10/05/2023 11:27 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 182.5 F 0.0  
2:00 183.0 F 2.9  
4:00 191.8 F 7.6  
6:00 211.8 F 13.3  
8:00 226.5 F 8.5  
10:00 235.2 F 11.1  
12:00 240.7 F 12.3  
14:00 241.5 F 11.6  
16:00 254.6 F 18.4  
18:00 266.5 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.0  
1:00 271.9 F 29.4  
2:00 271.4 F 29.5  
3:00 271.7 F 29.3  
4:00 271.3 F 29.5  
  
Min 271.0 F 28.9  
Max 272.4 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.1 F 0.0  
5:00 228.9 F 0.0  
10:00 221.5 F 0.0  
15:00 210.6 F 0.0  
20:00 204.7 F 0.0  
25:00 202.1 F 0.0  
30:00 202.3 F 0.0  
Drying Complete

FILLING: 1:51  
HEATING: 19:29  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 00:58:44  
CYCLE COMPLETE

10/05/2023 12:25 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_