

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1119

Name: RITTER 1  
Sterilizer ID: V2178198  
08/25/2021 11:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 141.4 F 0.0  
2:00 147.7 F 2.2  
4:00 174.9 F 5.8  
6:00 200.1 F 10.4  
8:00 215.0 F 4.9  
10:00 228.7 F 9.2  
12:00 236.5 F 10.3  
14:00 240.2 F 11.4  
16:00 237.7 F 9.5  
18:00 251.8 F 16.2  
20:00 263.3 F 22.9  
22:00 271.8 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.4  
1:00 272.6 F 28.9  
2:00 271.8 F 28.7  
3:00 272.4 F 28.9  
4:00 271.6 F 28.4

Min 271.5 F 28.3  
Max 272.9 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.2 F 0.0  
5:00 230.6 F 0.0  
10:00 203.0 F 0.0  
15:00 221.7 F 0.0  
20:00 227.2 F 0.0  
25:00 208.7 F 0.0  
30:00 204.0 F 0.0  
Drying Complete

FILLING: 1:18  
HEATING: 22:08  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:00:32  
CYCLE COMPLETE

08/25/2021 01:05 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02