

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
2.8uS / 1.7ppm  
CYCLE NUMBER 001392  
12:44 09/01/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:53  
PRESSURIZING 5:39  
STERILIZING 7:06  
136.0 C 220kPa 7:06  
136.4 C 221kPa 7:36  
136.2 C 220kPa 8:06  
136.2 C 220kPa 8:37  
135.8 C 217kPa 9:07  
136.4 C 221kPa 9:37  
136.4 C 222kPa 10:07  
135.9 C 218kPa 10:36  
VENTING 10:37  
AIR DRYING 11:31  
CYCLE COMPLETE 21:26  
-----

Digital Signature #  
ED630B205F4F98CB  
-----