

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1479

Name: RITTER 1  
Sterilizer ID: V2178198  
12/21/2021 08:22 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.0 F 0.0  
2:00 78.2 F 0.0  
4:00 109.4 F 0.0  
6:00 139.3 F 0.1  
8:00 155.8 F 0.4  
10:00 177.4 F 2.3  
12:00 194.6 F 4.3  
14:00 209.3 F 7.3  
16:00 220.4 F 10.6  
18:00 227.7 F 7.2  
20:00 231.0 F 8.8  
22:00 241.9 F 13.3  
24:00 239.8 F 10.6  
26:00 235.5 F 8.2  
28:00 249.5 F 14.8  
30:00 260.4 F 20.9  
32:00 270.4 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.1  
1:00 272.5 F 28.9  
2:00 272.1 F 28.3  
3:00 271.4 F 28.2  
4:00 272.6 F 29.0  
  
Min 271.3 F 28.0  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.4 F 0.0  
5:00 208.7 F 0.0  
10:00 223.0 F 0.0  
15:00 225.8 F 0.0  
20:00 230.6 F 0.0  
25:00 229.8 F 0.0  
30:00 226.3 F 0.0  
  
Drying Complete

FILLING: 1:01  
HEATING: 32:57  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:11:12  
CYCLE COMPLETE

12/21/2021 09:33 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_