

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3795

Name: RITTER2  
Sterilizer ID: V2313228  
07/17/2024 03:10 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 180.0 F 0.0  
2:00 180.6 F 2.8  
4:00 183.3 F 7.1  
6:00 199.1 F 12.1  
8:00 218.3 F 6.6  
10:00 229.2 F 11.5  
12:00 232.5 F 8.8  
14:00 238.9 F 10.8  
16:00 245.5 F 13.7  
18:00 257.3 F 20.1  
20:00 268.0 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 29.8  
1:00 272.5 F 30.2  
2:00 272.8 F 30.0  
3:00 272.5 F 29.8  
4:00 272.1 F 29.8  
  
Min 272.0 F 29.7  
Max 272.9 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.3 F 0.0  
5:00 219.6 F 0.0  
10:00 208.2 F 0.0  
15:00 191.8 F 0.0  
20:00 187.0 F 0.0  
25:00 187.7 F 0.0  
30:00 203.7 F 0.0  
Drying Complete

FILLING: 1:37  
HEATING: 21:53  
STERILIZING: 4:00  
VENTING: 3:48  
DRYING: 30:00  
TOTAL CYCLE: 01:01:18  
CYCLE COMPLETE

07/17/2024 04:11 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_