
Midmark M11 - v1.0.5
Total Cycles: 2027

Name: RITTER2
Sterilizer ID: V2313228
02/08/2023 02:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 206.6 F 0.0
2:00 195.4 F 3.5
4:00 202.5 F 8.4
6:00 218.2 F 9.4
8:00 232.0 F 8.5
10:00 239.1 F 10.8
12:00 245.0 F 13.2
14:00 245.1 F 12.9
16:00 258.8 F 20.2
18:00 270.7 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.5
1:00 272.8 F 29.4
2:00 271.9 F 28.5
3:00 272.7 F 29.4
4:00 272.4 F 28.8

Min 271.6 F 28.3
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.1 F 0.0
5:00 225.0 F 0.0
10:00 234.2 F 0.0
15:00 226.6 F 0.0
20:00 225.5 F 0.0
25:00 227.0 F 0.0
30:00 226.1 F 0.0
Drying Complete

FILLING: 1:58
HEATING: 18:38
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:57:40
CYCLE COMPLETE

02/08/2023 03:38 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____